

# MicroBioLogics®

## Mean Assay Value Information

The mean assay value provided on the Certificate of Assay label is the average assay value obtained by MicroBioLogics. Specific methods and materials are employed in determining the mean assay value. Your results may differ from the assay value listed on the product's label due to the use of different materials and methods. However, MicroBioLogics guarantees that if processed as directed, your results will be within the specific CFU concentration range designated for each individual product line. Please use the chart below for CFU concentration ranges.

Product	CFU
EZ-CFU™	<100 CFU per 0.1 mL
EZ-CFU™ One Step	<100 CFU per 0.1 mL
EZ-Accu Shot™	<100 CFU per 0.1 mL
EZ-PEC™	1.0E+05 to 1.0E+06 CFU per mL of product tested
EZ-FPC™	Qualitative: 1.0-9.9E+02 CFU per pellet Quantitative: 1.0-9.9E+03 CFU per pellet
EZ-SPORE™	1.0-9.9E+04 CFU per pellet
Epower™	E2: 1.0-9.9E+02 per pellet E3: 1.0-9.9E+03 per pellet E4: 1.0-9.9E+04 per pellet E6: 1.0-9.9E+06 per pellet E7: 1.0-9.9E+07 per pellet E8: 1.0-9.9E+08 per pellet